

## Expiry Extension: Yocon Virus Sampling Kit Viral Transport Media and Swabs

(February 1, 2022)

Effective immediately, Ontario Health has extended the expiry date for Yocon Virus Sampling Kit Viral Transport Media and Swabs (product number MT0301 – Oral, Nasal and Nasopharyngeal) from 1 year to 2 years post-date of manufacturing.

As this 2-year total shelf-life will not be incorporated into the product packaging, labelling, or instructions for use, please follow the below guide to determine the expiry date of your kit based on the lot number marked on the product:

Lot Number	Marked Expiry Date	Ontario Health Extension
01200701	2021-06-30	2022-06-30
01200703	2021-07-02	2022-07-02
01200706	2021-07-05	2022-07-05
01200709	2021-07-08	2022-07-08
01200710	2021-07-09	2022-07-09
01200909	2021-09-08	2022-09-08
01200910	2021-09-09	2022-09-09
01200913	2021-09-12	2022-09-12
01200915	2021-09-14	2022-09-14
01200916	2021-09-15	2022-09-15
01200918	2021-09-17	2022-09-17
01200919	2021-09-18	2022-09-18
01200920	2021-09-19	2022-09-19
01200922	2021-09-21	2022-09-21
01200923	2021-09-22	2022-09-22
01200924	2021-09-23	2022-09-23
01200927	2021-09-26	2022-09-26
01200928	2021-09-27	2022-09-27
01200929	2021-09-28	2022-09-28
01201012	2021-10-11	2022-10-11
01201018	2021-10-17	2022-10-17
01201023	2021-10-22	2022-10-22
01201026	2021-10-25	2022-10-25
01201103	2021-11-02	2022-11-02
01201105	2021-11-04	2022-11-04
01201116	2021-11-15	2022-11-15
01201117	2021-11-16	2022-11-16
01201118	2021-11-17	2022-11-17
01201119	2021-11-18	2022-11-18
01201120	2021-11-19	2022-11-19
01201123	2021-11-21	2022-11-21
	2021-11-23	2022-11-23
01201124	2021-11-23	2022-11-23
01201125	2021-11-24	2022-11-24
01201126	2021-11-25	2022-11-25
01201127	2021-11-26	2022-11-26
01201130	2021-11-29	2022-11-29

Inquiries and feedback may be directed to: [covid19testing@ontariohealth.ca](mailto:covid19testing@ontariohealth.ca)

Thank you,  
Ontario Health